

-----  
Midmark M11 - v1.0.0  
Total Cycles: 781

Name: RITTER 1  
Sterilizer ID: V2178198  
04/14/2021 07:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 72.4 F 0.0  
4:00 89.0 F 0.0  
6:00 125.6 F 0.1  
8:00 158.4 F 0.2  
10:00 186.6 F 1.5  
12:00 220.6 F 4.7  
14:00 224.6 F 8.0  
16:00 231.7 F 11.7  
18:00 228.8 F 6.2  
20:00 241.0 F 11.1  
22:00 234.3 F 7.9  
24:00 244.2 F 11.5  
26:00 245.8 F 12.8  
28:00 258.5 F 19.6  
30:00 269.5 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.3 F 28.3  
2:00 272.4 F 28.6  
3:00 271.6 F 27.8  
4:00 272.6 F 28.5

Min 271.5 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
5:00 237.0 F 0.0  
10:00 227.0 F 0.0  
15:00 223.6 F 0.0  
20:00 200.2 F 0.0  
25:00 222.5 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:58  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:08:43  
CYCLE COMPLETE

04/14/2021 08:12 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01