
Midmark M11 - v1.0.5
Total Cycles: 3778

Name: RITTER2
Sterilizer ID: V2313228
07/12/2024 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.3 F 0.0
2:00 156.2 F 2.7
4:00 182.3 F 7.1
6:00 205.7 F 12.7
8:00 223.9 F 6.9
10:00 238.7 F 12.3
12:00 236.7 F 9.7
14:00 236.6 F 8.8
16:00 249.6 F 15.3
18:00 262.1 F 22.5
20:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.3 F 29.3
2:00 271.3 F 28.9
3:00 272.3 F 29.3
4:00 271.2 F 28.7

Min 271.1 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 234.6 F 0.0
10:00 226.8 F 0.0
15:00 233.2 F 0.0
20:00 227.5 F 0.0
25:00 216.6 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

07/12/2024 04:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____