
Midmark M11 - v1.0.5
Total Cycles: 4166

Name: RITTER2
Sterilizer ID: V2313228
11/28/2024 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.1 F 0.0
2:00 87.6 F 0.0
4:00 101.2 F 0.1
6:00 141.4 F 0.4
8:00 168.1 F 1.5
10:00 191.2 F 5.0
12:00 223.4 F 8.6
14:00 230.8 F 12.9
16:00 232.0 F 8.4
18:00 243.1 F 13.5
20:00 240.9 F 11.5
22:00 239.1 F 10.4
24:00 252.5 F 17.0
26:00 264.0 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.2
2:00 272.1 F 29.5
3:00 272.2 F 28.9
4:00 272.6 F 29.7

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 231.2 F 0.0
10:00 222.8 F 0.0
15:00 184.2 F 0.0
20:00 174.7 F 0.0
25:00 174.9 F 0.0
30:00 185.0 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 27:57
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:06:24
CYCLE COMPLETE

11/28/2024 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____