
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001223
9:38 19/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:23
STERILIZING 4:38
136.0 C 220kPa 4:38
136.4 C 223kPa 5:08
136.4 C 225kPa 5:38
136.2 C 221kPa 6:08
136.2 C 223kPa 6:38
136.4 C 224kPa 7:08
136.4 C 222kPa 7:38
136.3 C 224kPa 8:08
136.3 C 224kPa 8:08
VENTING 8:08
AIR DRYING 9:00
CYCLE COMPLETE 24:00

Digital Signature #
107A0160FB22F6F6
