

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3695

Name: RITTER 1  
Sterilizer ID: V2178198  
08/29/2023 08:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 74.0 F 0.0  
4:00 85.2 F 0.0  
6:00 106.2 F 1.4  
8:00 136.5 F 3.7  
10:00 166.1 F 6.5  
12:00 186.1 F 9.5  
14:00 203.5 F 13.1  
16:00 209.4 F 5.0  
18:00 219.9 F 8.4  
20:00 231.2 F 12.3  
22:00 231.0 F 7.7  
24:00 242.0 F 12.6  
26:00 239.6 F 10.3  
28:00 251.6 F 16.1  
30:00 261.9 F 22.0  
32:00 271.6 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.2 F 28.7  
2:00 272.1 F 28.9  
3:00 271.9 F 28.4  
4:00 272.5 F 29.1  
  
Min 271.4 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
5:00 200.1 F 0.0  
10:00 208.0 F 0.0  
15:00 214.8 F 0.0  
20:00 221.4 F 0.0  
25:00 229.9 F 0.0  
30:00 229.2 F 0.0  
  
Drying Complete

FILLING: 1:01  
HEATING: 32:35  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:10:58  
CYCLE COMPLETE

08/29/2023 10:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_