

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4201

Name: RITTER2  
Sterilizer ID: V2313228  
12/10/2024 07:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 90.7 F 0.0  
2:00 94.6 F 0.0  
4:00 115.2 F 0.2  
6:00 155.2 F 3.7  
8:00 181.8 F 7.4  
10:00 209.9 F 11.8  
12:00 217.8 F 4.9  
14:00 227.1 F 8.9  
16:00 240.6 F 13.6  
18:00 239.0 F 10.8  
20:00 236.3 F 9.0  
22:00 250.3 F 15.8  
24:00 262.2 F 22.6  
26:00 272.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 272.0 F 28.9  
2:00 272.7 F 29.4  
3:00 272.1 F 29.4  
4:00 271.9 F 28.8

Min 271.6 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.9 F 0.0  
5:00 224.3 F 0.0  
10:00 225.6 F 0.0  
15:00 200.7 F 0.0  
20:00 185.5 F 0.0  
25:00 182.2 F 0.0  
30:00 184.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 26:15  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:04:35  
CYCLE COMPLETE

12/10/2024 08:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_