
Midmark M11 - v1.0.5
Total Cycles: 1031

Name: RITTER2
Sterilizer ID: V2313228
04/26/2022 06:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.4 F 0.0
2:00 75.7 F 0.0
4:00 82.0 F 0.0
6:00 105.8 F 0.0
8:00 160.2 F 0.3
10:00 202.5 F 2.3
12:00 207.7 F 5.8
14:00 210.9 F 9.4
16:00 222.0 F 13.1
18:00 229.7 F 7.9
20:00 241.2 F 12.4
22:00 238.4 F 9.9
24:00 237.8 F 9.2
26:00 249.1 F 14.9
28:00 260.8 F 21.5
30:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.6
1:00 272.4 F 29.3
2:00 271.5 F 28.8
3:00 272.3 F 28.8
4:00 272.1 F 29.4

Min 271.4 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 228.5 F 0.0
10:00 235.8 F 0.0
15:00 220.2 F 0.0
20:00 213.2 F 0.0
25:00 208.5 F 0.0
30:00 208.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:28
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

04/26/2022 08:19 AM
ACCEPTED: X
REJECTED:
Operator: U02