

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2083

Name: RITTER 1  
Sterilizer ID: V2178198  
07/04/2022 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 120.8 F 0.0  
2:00 133.8 F 0.2  
4:00 155.5 F 2.1  
6:00 180.7 F 5.0  
8:00 207.0 F 9.3  
10:00 218.0 F 9.1  
12:00 226.8 F 7.7  
14:00 237.7 F 12.4  
16:00 235.4 F 9.2  
18:00 238.8 F 9.6  
20:00 246.7 F 13.6  
22:00 259.0 F 20.2  
24:00 269.8 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 28.9  
2:00 271.4 F 28.3  
3:00 272.5 F 29.0  
4:00 271.6 F 28.3

Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 215.7 F 0.0  
10:00 219.0 F 0.0  
15:00 230.8 F 0.0  
20:00 250.1 F 0.0  
25:00 234.3 F 0.0  
30:00 215.2 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:57  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:02:58  
CYCLE COMPLETE

07/04/2022 03:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_