
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000257
10:50 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:49
STERILIZING 5:16
136.0 C 230kPa 5:16
136.4 C 233kPa 5:46
136.4 C 232kPa 6:16
136.4 C 233kPa 6:46
136.3 C 230kPa 7:16
136.3 C 231kPa 7:46
136.2 C 232kPa 8:16
136.2 C 232kPa 8:46
136.2 C 232kPa 8:46
VENTING 8:46
AIR DRYING 9:47
CYCLE COMPLETE 24:47

Digital Signature #
9A81022FD6FC7DD4
