
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008463
11:37 26/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:06
136.0 C 220kPa 3:06
136.4 C 225kPa 3:36
136.2 C 223kPa 4:06
136.2 C 221kPa 4:36
136.4 C 224kPa 5:06
136.1 C 222kPa 5:36
136.4 C 223kPa 6:06
136.1 C 223kPa 6:36
136.1 C 223kPa 6:36
VENTING 6:37
AIR DRYING 7:28
CYCLE COMPLETE 37:28

Digital Signature #
525B83630675A01E
