

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000353  
10:25 20/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:10  
PRESSURIZING 2:21  
STERILIZING 3:21  
136.0 C 220kPa 3:21  
136.3 C 230kPa 3:51  
136.2 C 231kPa 4:21  
136.3 C 232kPa 4:51  
136.4 C 231kPa 5:21  
136.4 C 233kPa 5:51  
136.2 C 232kPa 6:21  
136.3 C 232kPa 6:51  
136.3 C 232kPa 6:51  
VENTING 6:51  
AIR DRYING 7:43  
CYCLE COMPLETE 7:49  
-----

Digital Signature #  
FA0C81F6D8D6F157  
-----