

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000869  
12:48 14/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:07  
STERILIZING 3:12  
136.0 C 227kPa 3:12  
136.4 C 232kPa 3:42  
136.3 C 233kPa 4:12  
136.4 C 233kPa 4:42  
136.4 C 231kPa 5:12  
136.4 C 233kPa 5:42  
136.4 C 233kPa 6:12  
136.3 C 233kPa 6:42  
136.3 C 233kPa 6:42  
VENTING 6:42  
AIR DRYING 7:34  
CYCLE COMPLETE 22:34  
-----

Digital Signature #  
BA4C9F4BA556AC74  
-----