
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001012
13:22 17/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:40
136.4 C 225kPa 3:40
136.1 C 221kPa 4:11
136.4 C 224kPa 4:41
136.2 C 223kPa 5:11
136.4 C 224kPa 5:41
136.3 C 224kPa 6:11
136.2 C 221kPa 6:41
136.4 C 225kPa 7:11
136.4 C 225kPa 7:11
VENTING 7:11
AIR DRYING 8:02
CYCLE COMPLETE 9:30

Digital Signature #
680D47E050E72572
