
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002663
16:21 07/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:02
STERILIZING 3:02
136.4 C 226kPa 3:02
136.2 C 223kPa 3:32
136.0 C 222kPa 4:02
136.4 C 224kPa 4:32
136.4 C 224kPa 5:02
136.1 C 223kPa 5:32
136.4 C 223kPa 6:02
136.4 C 223kPa 6:32
136.4 C 223kPa 6:32
VENTING 6:32
AIR DRYING 7:19
CYCLE COMPLETE 17:10

Digital Signature #
9A112309FB2E3570
