
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002034
10:30 13/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:11
STERILIZING 3:13
136.0 C 222kPa 3:13
136.2 C 224kPa 3:43
135.9 C 221kPa 4:13
136.3 C 225kPa 4:43
136.4 C 226kPa 5:13
136.2 C 224kPa 5:43
136.4 C 227kPa 6:13
136.4 C 226kPa 6:43
136.4 C 226kPa 6:43
VENTING 6:43
AIR DRYING 7:34
CYCLE COMPLETE 16:42

Digital Signature #
B3303562589D6244
