

-----  
Midmark M11 - v1.0.5  
Total Cycles: 492

Name: RITTER2  
Sterilizer ID: V2313228  
10/27/2021 01:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.2 F 0.0  
2:00 147.3 F 2.3  
4:00 161.3 F 6.1  
6:00 187.9 F 11.2  
8:00 225.4 F 6.1  
10:00 232.3 F 11.5  
12:00 234.7 F 8.8  
14:00 238.0 F 10.4  
16:00 247.3 F 14.0  
18:00 259.5 F 20.7  
20:00 270.5 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 271.8 F 29.1  
2:00 272.3 F 29.3  
3:00 272.8 F 29.6  
4:00 271.9 F 29.8

Min 271.2 F 29.1  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 223.0 F 0.0  
10:00 206.2 F 0.0  
15:00 194.1 F 0.0  
20:00 190.1 F 0.0  
25:00 189.2 F 0.0  
30:00 197.5 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

10/27/2021 02:09 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01