
Midmark M11 - v1.0.0
Total Cycles: 5321

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2025 03:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.4 F 0.0
2:00 92.0 F 0.8
4:00 114.6 F 2.9
6:00 140.1 F 5.6
8:00 171.0 F 8.4
10:00 192.5 F 12.8
12:00 206.9 F 3.5
14:00 217.9 F 7.1
16:00 227.3 F 10.8
18:00 231.2 F 7.4
20:00 238.6 F 10.9
22:00 234.6 F 7.8
24:00 247.1 F 14.1
26:00 257.5 F 19.7
28:00 267.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.2 F 29.5
2:00 272.2 F 29.1
3:00 271.3 F 28.6
4:00 271.7 F 29.2

Min 271.2 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 195.4 F 0.0
10:00 197.9 F 0.0
15:00 200.8 F 0.0
20:00 197.0 F 0.0
25:00 196.4 F 0.0
30:00 198.1 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 29:36
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:08:16
CYCLE COMPLETE

02/06/2025 04:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____