

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001397  
7:39 09/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:32  
PRESSURIZING 6:03  
STERILIZING 7:30  
136.0 C 220kPa 7:30  
136.3 C 222kPa 8:01  
136.4 C 221kPa 8:31  
136.4 C 223kPa 9:01  
136.4 C 223kPa 9:31  
136.4 C 221kPa 10:01  
136.4 C 222kPa 10:31  
136.4 C 224kPa 11:00  
VENTING 11:01  
AIR DRYING 12:07  
CYCLE COMPLETE 12:15

-----  
Digital Signature #  
8AD23632BE1E3885  
-----