
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001599
9:31 11/19/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:21
PRESSURIZING 7:13
STERILIZING 8:53
136.0 C 220kPa 8:53
136.1 C 219kPa 9:24
135.9 C 220kPa 9:54
136.1 C 218kPa 10:24
136.4 C 220kPa 10:54
136.4 C 222kPa 11:24
135.9 C 219kPa 11:54
136.4 C 220kPa 12:24
VENTING 12:24
AIR DRYING 13:30
CYCLE COMPLETE 43:30

Digital Signature #
E66530499AF36404
