
Midmark M11 - v1.0.0
Total Cycles: 4230

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2024 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 74.6 F 0.2
4:00 101.5 F 0.8
6:00 111.2 F 2.5
8:00 123.4 F 4.9
10:00 147.8 F 7.2
12:00 187.8 F 10.3
14:00 202.1 F 12.0
16:00 215.1 F 5.2
18:00 224.1 F 8.2
20:00 227.8 F 11.6
22:00 229.3 F 5.9
24:00 236.2 F 10.4
26:00 239.5 F 8.8
28:00 244.1 F 12.4
30:00 254.6 F 17.9
32:00 264.1 F 23.5
34:00 272.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 271.9 F 29.2
2:00 271.8 F 28.7
3:00 271.6 F 29.1
4:00 272.3 F 29.0

Min 271.0 F 28.5
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 216.0 F 0.0
10:00 212.8 F 0.0
15:00 215.7 F 0.0
20:00 217.6 F 0.0
25:00 220.0 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 34:12
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:12:45
CYCLE COMPLETE

01/25/2024 10:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____