
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005150
10:33 27/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:01
STERILIZING 2:57
136.2 C 220kPa 2:57
136.4 C 223kPa 3:27
136.4 C 223kPa 3:57
136.4 C 224kPa 4:27
136.4 C 222kPa 4:57
136.4 C 222kPa 5:27
136.4 C 223kPa 5:57
136.1 C 222kPa 6:27
136.1 C 222kPa 6:27
VENTING 6:27
AIR DRYING 7:15
CYCLE COMPLETE 10:25

Digital Signature #
6BBDC62CBE6864FD
