

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002044  
10:41 16/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:22  
STERILIZING 3:29  
136.5 C 226kPa 3:29  
136.3 C 225kPa 3:59  
136.4 C 225kPa 4:29  
136.3 C 226kPa 4:59  
136.4 C 226kPa 5:29  
136.4 C 226kPa 5:59  
136.4 C 226kPa 6:29  
136.2 C 225kPa 6:59  
136.2 C 225kPa 6:59  
VENTING 6:59  
AIR DRYING 7:50  
CYCLE COMPLETE 8:08  
-----

Digital Signature #  
9B0576884F4DCA79  
-----