
Midmark M11 - v1.0.5
Total Cycles: 3752

Name: RITTER2
Sterilizer ID: V2313228
07/03/2024 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.4 F 0.0
2:00 105.6 F 0.6
4:00 121.8 F 2.7
6:00 149.4 F 5.6
8:00 172.8 F 9.1
10:00 204.6 F 12.8
12:00 212.8 F 6.6
14:00 223.1 F 10.8
16:00 225.3 F 7.1
18:00 237.7 F 11.7
20:00 235.6 F 9.5
22:00 248.1 F 15.4
24:00 258.9 F 21.4
26:00 268.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.7
1:00 272.2 F 29.9
2:00 271.5 F 30.4
3:00 272.1 F 30.5
4:00 272.1 F 30.5

Min 270.9 F 29.5
Max 273.9 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 205.7 F 0.0
10:00 210.3 F 0.0
15:00 190.3 F 0.0
20:00 187.1 F 0.0
25:00 188.5 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:06
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:05:37
CYCLE COMPLETE

07/03/2024 01:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____