
Midmark M11 - v1.0.5
Total Cycles: 3268

Name: RITTER2
Sterilizer ID: V2313228
01/03/2024 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 81.7 F 0.0
4:00 119.3 F 0.1
6:00 100.6 F 1.2
8:00 132.1 F 4.0
10:00 161.2 F 7.2
12:00 198.3 F 10.6
14:00 217.2 F 4.8
16:00 227.2 F 7.6
18:00 234.4 F 11.7
20:00 230.6 F 7.5
22:00 241.8 F 12.7
24:00 239.7 F 10.7
26:00 252.0 F 16.7
28:00 262.4 F 22.9
30:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.5 F 29.4
2:00 272.6 F 29.9
3:00 272.2 F 29.7
4:00 272.4 F 29.3

Min 271.8 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 230.6 F 0.0
10:00 229.3 F 0.0
15:00 223.1 F 0.0
20:00 223.9 F 0.0
25:00 228.6 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:04
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

01/03/2024 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____