
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001617
11:01 11/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:11
PRESSURIZING 7:01
STERILIZING 8:41
136.5 C 222kPa 8:41
136.5 C 222kPa 9:11
136.3 C 218kPa 9:41
136.3 C 221kPa 10:12
135.9 C 220kPa 10:42
136.5 C 223kPa 11:12
136.3 C 222kPa 11:42
136.4 C 223kPa 12:12
VENTING 12:12
AIR DRYING 13:14
CYCLE COMPLETE 13:44

Digital Signature #
CCC6667777E136D5
