
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000251
6:56 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:00
PRESSURIZING 4:03
STERILIZING 5:44
136.0 C 229kPa 5:44
136.0 C 229kPa 6:14
136.3 C 231kPa 6:44
136.3 C 231kPa 7:14
136.4 C 232kPa 7:44
136.4 C 233kPa 8:14
136.3 C 232kPa 8:44
136.2 C 230kPa 9:14
136.2 C 230kPa 9:14
VENTING 9:14
AIR DRYING 10:19
CYCLE COMPLETE 25:19

Digital Signature #
9DDDE3F9F51B876D
