
Midmark M11 - v1.0.5
Total Cycles: 811

Name: RITTER2
Sterilizer ID: V2313228
02/16/2022 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.7 F 0.0
2:00 179.4 F 3.9
4:00 202.6 F 9.4
6:00 222.8 F 7.0
8:00 235.7 F 11.8
10:00 237.4 F 10.1
12:00 236.5 F 9.2
14:00 252.0 F 16.6
16:00 264.2 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 272.6 F 29.2
2:00 271.8 F 29.4
3:00 272.1 F 29.1
4:00 271.5 F 29.4

Min 271.0 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 235.1 F 0.0
10:00 192.8 F 0.0
15:00 185.1 F 0.0
20:00 182.8 F 0.0
25:00 182.1 F 0.0
30:00 181.4 F 0.0

Drying Complete

FILLING: 2:43
HEATING: 17:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

02/17/2022 07:44 AM
ACCEPTED: X
REJECTED:
Operator: U02