
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002054
12:45 15/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:27
PRESSURIZING 1:31
STERILIZING 2:27
136.0 C 220kPa 2:27
136.0 C 222kPa 2:57
136.3 C 221kPa 3:27
136.2 C 218kPa 3:57
136.4 C 224kPa 4:27
136.4 C 224kPa 4:57
136.4 C 223kPa 5:27
136.4 C 223kPa 5:57
136.4 C 223kPa 5:57
VENTING 5:57
AIR DRYING 6:50
CYCLE COMPLETE 6:59

Digital Signature #
B4546B50F5A64B75
