
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001967
11:36 16/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 2:01
STERILIZING 3:00
136.4 C 224kPa 3:00
136.1 C 224kPa 3:30
136.4 C 226kPa 4:00
136.2 C 223kPa 4:30
136.4 C 226kPa 5:00
136.3 C 225kPa 5:30
136.1 C 224kPa 6:00
136.4 C 226kPa 6:30
136.4 C 226kPa 6:30
VENTING 6:30
AIR DRYING 7:22
CYCLE COMPLETE 14:55

Digital Signature #
026B22ED1BB8C1A0
