
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004483
13:16 04/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:16
STERILIZING 4:16
136.0 C 220kPa 4:16
136.3 C 224kPa 4:46
136.3 C 224kPa 5:16
136.3 C 224kPa 5:46
136.1 C 223kPa 6:16
136.4 C 224kPa 6:46
136.4 C 225kPa 7:16
136.2 C 224kPa 7:46
136.2 C 224kPa 7:46
VENTING 7:46
AIR DRYING 8:37
CYCLE COMPLETE 8:45

Digital Signature #
0235C990BEEA555A
