
Midmark M11 - v1.0.5
Total Cycles: 263

Name: RITTER2
Sterilizer ID: V2313228
08/04/2021 01:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.5 F 0.0
2:00 93.7 F 0.6
4:00 101.2 F 2.3
6:00 126.7 F 5.2
8:00 158.2 F 8.8
10:00 183.0 F 12.7
12:00 212.2 F 6.3
14:00 213.9 F 10.4
16:00 231.6 F 8.4
18:00 232.0 F 11.9
20:00 238.5 F 9.7
22:00 249.6 F 15.4
24:00 260.3 F 21.3
26:00 269.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 273.6 F 30.2
1:00 272.1 F 30.2
2:00 272.2 F 30.4
3:00 271.9 F 30.4
4:00 270.6 F 30.4

Min 270.6 F 30.1
Max 273.6 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 229.6 F 0.0
10:00 219.0 F 0.0
15:00 231.7 F 0.0
20:00 171.7 F 0.0
25:00 156.8 F 0.0
30:00 148.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 27:30
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:05:40
CYCLE COMPLETE

08/04/2021 02:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____