
Midmark M11 - v1.0.0
Total Cycles: 5431

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2025 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 195.3 F 3.6
4:00 199.9 F 8.5
6:00 208.4 F 9.5
8:00 225.3 F 8.1
10:00 239.2 F 13.1
12:00 239.2 F 11.2
14:00 238.6 F 9.9
16:00 252.7 F 17.0
18:00 265.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.2
2:00 271.8 F 28.7
3:00 272.7 F 29.2
4:00 271.8 F 28.5

Min 271.5 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 200.3 F 0.0
10:00 224.2 F 0.0
15:00 231.3 F 0.0
20:00 225.1 F 0.0
25:00 221.6 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:42
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

03/21/2025 01:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____