
Midmark M11 - v1.0.5
Total Cycles: 1795

Name: RITTER2
Sterilizer ID: V2313228
11/30/2022 03:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 194.8 F 2.0
4:00 205.4 F 7.5
6:00 228.1 F 13.2
8:00 231.5 F 10.4
10:00 232.7 F 8.6
12:00 238.3 F 10.0
14:00 248.2 F 15.1
16:00 260.8 F 22.3
18:00 272.4 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.4 F 29.8
2:00 271.8 F 29.8
3:00 272.3 F 29.8
4:00 271.6 F 29.5

Min 271.5 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 230.5 F 0.0
10:00 239.3 F 0.0
15:00 238.7 F 0.0
20:00 227.7 F 0.0
25:00 232.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 18:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

11/30/2022 04:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____