
Midmark M11 - v1.0.5
Total Cycles: 2613

Name: RITTER2
Sterilizer ID: V2313228
07/05/2023 04:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.1 F 0.0
2:00 131.1 F 0.0
4:00 138.3 F 0.1
6:00 175.9 F 3.8
8:00 218.4 F 8.1
10:00 227.0 F 12.7
12:00 228.8 F 7.8
14:00 239.9 F 12.8
16:00 236.4 F 10.2
18:00 236.6 F 8.6
20:00 248.9 F 15.3
22:00 261.3 F 22.2
24:00 272.1 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.4 F 30.0
2:00 271.8 F 29.3
3:00 271.3 F 29.4
4:00 272.6 F 30.0

Min 271.1 F 29.0
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 210.6 F 0.0
10:00 213.7 F 0.0
15:00 212.6 F 0.0
20:00 211.0 F 0.0
25:00 210.2 F 0.0
30:00 218.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

07/05/2023 06:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____