
Midmark M11 - v1.0.5
Total Cycles: 2984

Name: RITTER2
Sterilizer ID: V2313228
10/06/2023 10:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.1
2:00 76.6 F 0.1
4:00 103.4 F 0.7
6:00 141.7 F 3.6
8:00 172.5 F 7.6
10:00 196.4 F 12.1
12:00 208.7 F 4.1
14:00 221.2 F 8.1
16:00 234.5 F 12.5
18:00 234.6 F 8.8
20:00 239.7 F 10.2
22:00 245.9 F 13.4
24:00 259.1 F 20.6
26:00 270.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 271.7 F 28.6
2:00 272.0 F 28.7
3:00 272.3 F 29.0
4:00 272.8 F 29.3

Min 271.6 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.7 F 0.0
5:00 211.5 F 0.0
10:00 199.8 F 0.0
15:00 197.1 F 0.0
20:00 194.5 F 0.0
25:00 192.2 F 0.0
30:00 187.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 26:38
STERILIZING: 4:00
VENTING: 2:45
DRYING: 30:00
TOTAL CYCLE: 01:04:58
CYCLE COMPLETE

10/06/2023 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____