
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002603
9:32 25/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:11
STERILIZING 4:19
136.4 C 224kPa 4:19
136.2 C 222kPa 4:49
136.0 C 223kPa 5:19
136.3 C 224kPa 5:49
136.4 C 225kPa 6:19
136.3 C 223kPa 6:49
136.2 C 224kPa 7:19
136.3 C 224kPa 7:49
136.3 C 224kPa 7:49
VENTING 7:49
AIR DRYING 8:42
CYCLE COMPLETE 10:51

Digital Signature #
7608116AE6735C1D
