
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000779
7:06 10/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:22
STERILIZING 4:21
136.5 C 221kPa 4:21
136.2 C 218kPa 4:51
136.3 C 221kPa 5:21
136.4 C 218kPa 5:51
136.1 C 220kPa 6:21
136.1 C 217kPa 6:51
136.2 C 221kPa 7:21
136.1 C 217kPa 7:51
136.1 C 217kPa 7:51
VENTING 7:51
AIR DRYING 8:46
CYCLE COMPLETE 9:15

Digital Signature #
D9216CA41C10A3D1
