

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002562  
8:59 07/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:47  
STERILIZING 3:50  
136.0 C 220kPa 3:50  
136.3 C 223kPa 4:20  
136.1 C 222kPa 4:50  
136.2 C 223kPa 5:20  
136.4 C 224kPa 5:50  
136.4 C 224kPa 6:20  
136.4 C 224kPa 6:50  
136.2 C 220kPa 7:20  
136.2 C 220kPa 7:20  
VENTING 7:20  
AIR DRYING 8:17  
CYCLE COMPLETE 9:09  
-----

Digital Signature #  
4CE2EDC441E91473  
-----