

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3282

Name: RITTER2  
Sterilizer ID: V2313228  
01/10/2024 11:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
2:00 178.5 F 2.5  
4:00 191.0 F 7.2  
6:00 209.2 F 12.1  
8:00 226.0 F 7.8  
10:00 235.5 F 13.1  
12:00 237.5 F 11.3  
14:00 238.1 F 10.4  
16:00 251.7 F 17.1  
18:00 263.1 F 23.9  
20:00 273.1 F 30.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.7  
1:00 272.6 F 30.8  
2:00 271.7 F 30.1  
3:00 272.8 F 30.5  
4:00 272.3 F 30.8

Min 270.8 F 29.5  
Max 273.3 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.0 F 0.0  
5:00 245.5 F 0.0  
10:00 233.3 F 0.0  
15:00 233.8 F 0.0  
20:00 218.8 F 0.0  
25:00 217.2 F 0.0  
30:00 210.1 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 20:17  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 00:59:45  
CYCLE COMPLETE

01/10/2024 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_