
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005402
10:16 21/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:54
STERILIZING 2:52
136.5 C 220kPa 2:52
136.3 C 222kPa 3:22
136.3 C 221kPa 3:52
136.3 C 222kPa 4:22
135.9 C 220kPa 4:52
136.4 C 223kPa 5:22
136.4 C 222kPa 5:52
136.4 C 223kPa 6:22
136.4 C 223kPa 6:22
VENTING 6:23
AIR DRYING 7:15
CYCLE COMPLETE 9:02

Digital Signature #
9CA21B7A92AA054E
