
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001412
8:39 09/09/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 5:39
STERILIZING 7:04
136.0 C 220kPa 7:04
136.2 C 222kPa 7:34
136.1 C 220kPa 8:04
136.4 C 223kPa 8:34
136.3 C 222kPa 9:04
136.4 C 223kPa 9:34
136.3 C 222kPa 10:04
136.3 C 222kPa 10:34
VENTING 10:34
AIR DRYING 11:38
CYCLE COMPLETE 11:56

Digital Signature #
C21D0FB19D41C6E3
