
Midmark M11 - v1.0.5
Total Cycles: 3829

Name: RITTER2
Sterilizer ID: V2313228
08/02/2024 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.4 F 0.0
2:00 99.7 F 0.1
4:00 135.0 F 0.2
6:00 166.2 F 3.4
8:00 191.7 F 7.9
10:00 211.6 F 12.6
12:00 217.4 F 5.5
14:00 230.1 F 9.8
16:00 232.8 F 8.7
18:00 239.9 F 11.4
20:00 236.4 F 9.2
22:00 251.6 F 16.4
24:00 264.1 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.1 F 29.2
2:00 272.8 F 29.6
3:00 272.6 F 29.1
4:00 272.1 F 28.7

Min 271.7 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.8 F 0.0
5:00 223.8 F 0.0
10:00 215.4 F 0.0
15:00 212.7 F 0.0
20:00 206.7 F 0.0
25:00 203.8 F 0.0
30:00 199.4 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 25:46
STERILIZING: 4:00
VENTING: 2:31
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

08/02/2024 01:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____