
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003528
11:44 26/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:45
STERILIZING 3:56
136.0 C 220kPa 3:56
136.2 C 223kPa 4:26
CYCLE ABORTED 4:40
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:40

Digital Signature #
930C93C084736EAF
