
Midmark M11 - v1.0.0
Total Cycles: 5198

Name: RITTER 1
Sterilizer ID: V2178198
12/17/2024 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.3 F 0.0
2:00 117.6 F 0.0
4:00 148.5 F 0.0
6:00 180.4 F 0.0
8:00 199.7 F 0.2
10:00 215.4 F 2.5
12:00 225.5 F 5.9
14:00 233.4 F 9.5
16:00 240.6 F 11.8
18:00 240.4 F 10.7
20:00 236.2 F 8.3
22:00 242.7 F 11.0
24:00 247.4 F 13.6
26:00 259.0 F 19.9
28:00 269.1 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.3 F 28.6
2:00 272.7 F 28.8
3:00 272.4 F 28.4
4:00 271.9 F 28.1

Min 271.6 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 213.7 F 0.0
10:00 227.6 F 0.0
15:00 224.5 F 0.0
20:00 226.9 F 0.0
25:00 211.9 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 29:16
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

12/17/2024 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____