
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 008191
13:13 02/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:54
STERILIZING 4:02
136.4 C 224kPa 4:02
136.4 C 222kPa 4:32
136.2 C 222kPa 5:02
136.4 C 221kPa 5:32
136.4 C 220kPa 6:02
136.4 C 223kPa 6:32
136.2 C 220kPa 7:02
136.4 C 224kPa 7:32
136.4 C 224kPa 7:32
VENTING 7:32
AIR DRYING 8:20
CYCLE COMPLETE 8:37

Digital Signature #
CAC4C5B90E54187D
