

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3442

Name: RITTER2  
Sterilizer ID: V2313228  
03/07/2024 05:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.3 F 0.0  
2:00 168.3 F 2.6  
4:00 182.1 F 6.9  
6:00 202.1 F 11.8  
8:00 220.7 F 6.6  
10:00 230.6 F 11.8  
12:00 234.4 F 9.5  
14:00 237.2 F 9.6  
16:00 248.5 F 14.9  
18:00 260.3 F 21.5  
20:00 271.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.6  
1:00 273.1 F 30.0  
2:00 271.7 F 29.7  
3:00 273.0 F 30.1  
4:00 271.6 F 29.6  
  
Min 271.6 F 29.5  
Max 273.2 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.2 F 0.0  
5:00 233.8 F 0.0  
10:00 223.9 F 0.0  
15:00 219.3 F 0.0  
20:00 216.7 F 0.0  
25:00 216.8 F 0.0  
30:00 217.6 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 21:36  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:00:53  
CYCLE COMPLETE

03/07/2024 06:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_