
Midmark M11 - v1.0.0
Total Cycles: 1115

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2021 06:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 85.2 F 0.0
4:00 86.2 F 0.0
6:00 123.7 F 0.0
8:00 149.3 F 0.1
10:00 180.1 F 0.2
12:00 202.0 F 1.8
14:00 216.6 F 4.8
16:00 225.2 F 7.4
18:00 233.6 F 10.8
20:00 231.9 F 8.0
22:00 241.1 F 11.5
24:00 236.9 F 9.1
26:00 239.8 F 9.7
28:00 248.1 F 14.1
30:00 260.6 F 20.9
32:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.7 F 28.8
2:00 271.6 F 27.9
3:00 272.7 F 28.7
4:00 271.6 F 27.9

Min 271.6 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 201.4 F 0.0
10:00 184.4 F 0.0
15:00 183.5 F 0.0
20:00 179.8 F 0.0
25:00 179.5 F 0.0
30:00 182.3 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 32:28
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:10:26
CYCLE COMPLETE

08/25/2021 08:34 AM
ACCEPTED: X
REJECTED:
Operator: U02