
Midmark M11 - v1.0.0
Total Cycles: 1697

Name: RITTER 1
Sterilizer ID: V2178198
03/04/2022 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 73.9 F 0.0
4:00 89.2 F 0.2
6:00 124.2 F 0.9
8:00 151.0 F 2.3
10:00 175.8 F 4.1
12:00 193.9 F 6.7
14:00 206.2 F 10.3
16:00 216.4 F 7.9
18:00 225.6 F 7.1
20:00 236.0 F 11.3
22:00 233.4 F 7.8
24:00 246.1 F 13.0
26:00 244.4 F 12.2
28:00 256.8 F 18.6
30:00 267.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 271.8 F 27.9
2:00 272.4 F 28.3
3:00 272.5 F 28.7
4:00 271.4 F 27.9

Min 271.3 F 27.7
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 209.7 F 0.0
10:00 209.5 F 0.0
15:00 205.2 F 0.0
20:00 210.6 F 0.0
25:00 211.4 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

03/04/2022 03:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____