
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006004
7:35 19/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:59
STERILIZING 3:58
136.0 C 220kPa 3:58
135.7 C 219kPa 4:28
135.9 C 220kPa 4:58
136.2 C 221kPa 5:28
136.0 C 220kPa 5:58
136.4 C 223kPa 6:28
136.3 C 218kPa 6:58
136.4 C 222kPa 7:28
136.4 C 222kPa 7:28
VENTING 7:28
AIR DRYING 8:18
CYCLE COMPLETE 9:05

Digital Signature #
3093EB26991BC440
