
Midmark M11 - v1.0.0
Total Cycles: 4906

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2024 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 190.3 F 0.2
4:00 193.5 F 0.8
6:00 211.4 F 4.3
8:00 223.0 F 8.3
10:00 236.1 F 12.5
12:00 233.1 F 7.9
14:00 244.3 F 13.0
16:00 240.7 F 10.6
18:00 236.5 F 8.6
20:00 250.7 F 15.2
22:00 261.6 F 21.4
24:00 271.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 271.5 F 28.1
2:00 272.6 F 28.7
3:00 272.3 F 28.3
4:00 271.4 F 27.9

Min 271.2 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 210.3 F 0.0
10:00 233.2 F 0.0
15:00 231.3 F 0.0
20:00 228.3 F 0.0
25:00 217.1 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:38
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:15
CYCLE COMPLETE

09/13/2024 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____